

Return & Repair Form:

Customer Information:

Company Name: New Jersey State Police B: _____ S: _____
Date Received: 1-16-09 Date Given to Service: _____
Carrier: FedEx UPS DHL - USPS Method: GRD NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 7110 6510 - 6810 Serial #: AR 105-0004
Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
Whole Instrument Top ½ Sim Ser #: _____
Other _____ Probe Ser # DD P -
ACCESSORIES:
☐ 110 V A/C Adapter ☐ Regulator ☐ Mag Card Rdr# _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: ERROR 112

Part Number	Description	Qty	Total Cost
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	

Repair Notes:

CLEARED ERROR 112 / FC 15 "WE" = MAY '05 /
CAL w/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 01-20-09