

# RETURN AND REPAIR FORM

**Dräger**

## Customer Information

B:

S:

Company Name: Peapack - Gladstone P.D., Peapack, NJ

Date Received: 09/20/2016

Date given to service: 09/20/2016

Carrier: ☐ FedEx ☒ UPS ☐ USPS Shipping Method: ☒ GRD ☐ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR WJ-0005

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| 4414161     | Calibration | 1   | NC/W       |
| 4414166     | Labor       | 5   | NC/W       |
| 8315075     | Motherboard | 1   | NC/W       |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |

Repair Notes: Replaced defective motherboard.

Performed Linearity Test and passed.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 12/05/2016