

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Barnegat Twp P.D. , Barnegat NJ

Date Received: 05/22/12

Date given to service: 05/22/12

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☐ GRD ☐ NDA-PRI

☒ 3DAY ☐ 2DAY
☐ NDA-STD

Product: ☐ 6510

☐ 6810

☒ 7110

☐ 8610

☐ 7510

☐ DT5000

☐ 7410 Upper-half

☐ 9510

☐ 7410 Whole

Serial Number: AR WJ-0007

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Regulator

☐ Printer Ribbon

☐ Printer Paper

☐ Mouthpieces

☐ 9510 Stylus

☐ 9510 Top Cover

☐ Carrying Case

☐ Dry Gas

☐ Other (please specify) _____

Repair Information

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	.5	NC/W

Repair Notes: Replaced capacitor at location C-156 to restore modem connection.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 05-24-12