

Return & Repair Form

Customer Information:

Company Name: New Jersey State PD B: _____ S: _____

Date Received: 7-18-08 Date Given to Service: _____

Carrier: FedEx UPS DHL - USPS Method: GRD NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810

Serial #: AR 123-0009

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: _____

Other _____

Probe Ser # DD _____ P -

ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) _____

Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: NO POWER + E72

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>192.00</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>120-</u>
<u>MPLABOR</u>		<u>.5</u>	<u>44.00</u> → +S/H/I

Repair Notes:

POWER SUPPLY COVER LOOSE / WIRE LOOSE / NOISE ERROR 72 NTC FOR BREATH HOSE
INSULATION FOR HEATER WIRE PULLED OUT - REPLACE BREATH HOSE
HOSE HEAT WIRE HARNESS LOOSE - CAUSED E72 - CORRECTED
REPLACE FC
CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 07-29-08