

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Surf City P.D. Surf City, N.J.

Date Received: 09/07/11 Date given to service: _____

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WJ-00012

Printer Serial#: AR _____

Sim Serial#: DD _____

Probe Serial#: DD _____

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
6808455	Fuel Cell	1	N/C W
6808486	Plate	1	N/C W
mp Cal 71	Calibration	1	N/C W
mp Labor	Labor	.5	N/C W

Repair Notes: Replaced Fuel Cell ARWF-0115, with Fuel Cell ARCH-1199, also replaced check valve because of white particles on front and back of check valve.
Cal w/QC & ops check

Service Technician: [Signature]

Date: 9-9-11