

Return & Repair Form

Customer Information:

Company Name: Point Pleasant Borough Police Dept. B: _____ S: _____

Date Received: 4-3-08 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 7110 - 6510 - 6810

Serial #: AR WJ -0014

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top ½

Sim Ser #: _____

Other _____

Probe Ser # DD P -

ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>MPMISC</u>	<u>IR DETECT. PCB PN 680945B (NEW STOCK)</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL 71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACE IR DETECT. PCB / CAL w/ QC & OPS CHECKS

Service Technician: RE

Date: 04-04-08