

# Return & Repair Form:

## Customer Information:

Company Name: Pain + Pleasant Borough B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 7-22-09 Date Given to Service: \_\_\_\_\_

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810

Serial #: ARWJ - 0014

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_

Probe Ser # DD P -

## ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# \_\_\_\_\_

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # \_\_\_\_\_

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) \_\_\_\_\_

Warranty Exp. Date \_\_\_\_\_

## Repair Information:

NEW Cuvette

Test #

Reason for Return:

| Part Number      | Description                | Qty       | Total Cost      |
|------------------|----------------------------|-----------|-----------------|
| <u>10809500</u>  | <u>cuvette</u>             | <u>1</u>  | <u>N/C WARR</u> |
| <u>10809512</u>  | <u>Grommit with Ventil</u> | <u>1</u>  | <u>N/C WARR</u> |
| <u>MP Cal 71</u> | <u>Cal</u>                 | <u>1</u>  | <u>N/C WARR</u> |
|                  |                            |           |                 |
|                  |                            |           |                 |
|                  |                            |           |                 |
|                  |                            |           |                 |
| <u>MP Labor</u>  | <u>Labor</u>               | <u>.5</u> | <u>N/C WARR</u> |

Repair Notes:

Re place Cuvette

Cal w/Dec Ops Check

Service Technician: JS

Date: 7/23/2009