

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Middletown Twp. P.D., Middletown, NJ

Date Received: 08/28/13

Date given to service: 08/28/13

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☒ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WJ-0015

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: November 2013

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter
☐ Mouthpieces
☐ Dry Gas

☐ Regulator
☐ 9510 Stylus
☐ Other (please specify) _____

☐ Printer Ribbon
☐ 9510 Top Cover

☐ Printer Paper
☐ Carrying Case

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Customer reported Error 112 and was able to clear it.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 08/29/13