

# RETURN AND REPAIR FORM

# Dräger

## Customer Information

B:

S:

Company Name: Middletown Twp. P.D., Middletown, NJ

Date Received: 05/06/15

Date given to service: 05/06/15

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☒ GRD ☐ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR WJ-0015

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: November 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

| Part Number | Description | Qty | Total Cost |
|-------------|-------------|-----|------------|
| 4414161     | Calibration | 1   | NC/W       |
| 4414166     | Labor       | .5  | NC/W       |
| 6808455     | Fuel Cell   | 1   | NC/W       |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |
|             |             |     |            |

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARFM-1971

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 05/18/15