

Dräger safety

Internal Use

Return & Repair Form

Customer Information:

CUSTOMER: Seaside Hqts PD

Cust # 60725/60729

Return: 9/6/07

Carrier: Fed

Method: Priority Overnight

Product Information:

Model: 7110 7410 5510 6810

Serial # ARWJ-0018

Description: A B PLUS SCREENER

Printer Serial #

Whole Inst. Top Half

Simulator Serial #

Probe Serial #

Unit ☐ Returned to stock on:

Accessories: (check all that apply):

☒ 110VAC Adapter

☐ Regulator

☐ Mag Card Rdr #

☐ Printer Paper

☐ Printer Ribbon

☐ Casio #

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (specify):

Warranty Expiration Date:

Test Information:

Test #

Reason for Return: LOW FC RESULTS

Part #	Description	Qty	Total
6808455	FC	1	WARR
MPLAC71	7110 CAL	1	WARR
MPLABOR	LABOR	.5	WARR

Repair Notes: REPLACED F/C / CAL w/ QC: OPS CHECKS

Technician:

Date: 09.06.07