

Return & Repair Form

Customer Information:

CUSTOMER Northfield P.D. Cust # _____
Date of Return: 5/21/07 Carrier: FedEx Method: Std.

Product Information:

Product: 7110 7410 6510 6810 Serial # ARUM-0022

Description: A B PLUS SCREENER Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): POWER CORD

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: EC problem

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced questionable F/C. Full cal ops
check + Q.C.

Technician: Q40

Date: 5/21/07