

# Return & Repair Form

## Customer Information:

Company Name: Worcester Police Dept. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 7-14-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR WM - 0022  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Reason for Return: STUCK ON "SEARCHING" + ERR 71 Test # \_\_\_\_\_

| Part Number     | Description | Qty      | Total Cost      |
|-----------------|-------------|----------|-----------------|
| <u>MPCAL 71</u> | <u>CAL</u>  | <u>1</u> | <u>N/C WARR</u> |
|                 |             |          |                 |
|                 |             |          |                 |
|                 |             |          |                 |
|                 |             |          |                 |
|                 |             |          |                 |
|                 |             |          |                 |

## Repair Notes:

FCIS YA SERIES (INSTALLED 5-21-07) / CAL w/ QC & OPS CHECKS  
PROBLEM NOT DUPLICATED

Service Technician: [Signature]

Date: 07-25-08