Return & Repair Form:

Customer Information:

Company Name: NORTHFIELD POLICE DEPT B: _____ S: _____

Date Received: 01-05-09 Date Given to Service: 01-05-09

Carrier: FedEx – UPS – DHL – USPS **Method:** GRD – NDA – 2DY – 3DY – Other

Product Information:

Product: 7410 – 7110 – 6510 – 6810 Serial #: AR WM - 0022

Description: A – B – Plus – Screener – Demo Printer Ser #: AR _____ - _____

Whole Instrument **Top ½** Sim Ser #: _____

Other _____ Probe Ser # DD _____ P - _____

ACCESSORIES:

☐ 110 V A/C Adapter ☐ Regulator ☐ Mag Card Rdr# _____

☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____

☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Reason for Return: 1/16/11 IR RESULTS **Test #**

Part Number	Description	Qty	Total Cost
6809500	CUVETTE	1	N/C WARR
MPCAL71	CAL	1	N/C WARR
MISC LABOR		1 HR	N/C WARR

Repair Notes:

FUEL CELL WAS INSTALLED 5-21-07 (YA SERIES) / CUVETTE HAS SPOTS ON MIRRORS / REPLACE CUVETTE / CAL W/ QC & OPS CHECKS

Service Technician: _____ Date: 01-06-09