

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Northfield P.D. Northfield N.J.

Date Received: 09/20/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WM-0022
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel cell</u>	<u>1</u>	<u>N/C W</u>
<u>MP CAL 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>mp labor</u>	<u>Labor</u>	<u>-5</u>	<u>N/C W</u>

Repair Notes: Fuel cell ARYA-0308 replaced because of leakage, replaced with Fuel cell ARCH-1179.

Cal W/RC: ops check

Service Technician: [Signature] Date: 9/28/11