

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Northfield P.D., Northfield NJ

Date Received: 04-06-12

Date given to service: 04-06-12

Carrier:  FedEx  UPS  USPS

Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WM-0022

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
8315075	PCB Board	1	N/C W
MP CAL 71	Calibration	1	N/C W
MP Labor	Labor	.5	N/C W

Repair Notes: While the customer was performing a part two linearity test the instrument displayed and reported a different concentration then the 0.080 that was entered. CAL / W QC AND OPS CHECK

Service Technician JS

Date: 04-24-12