

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Clifton Police Dept. Cust # 59806
 Date of Return: 11/20/06 Carrier: Fed Ex Method: Priority

Product Information:

Product : (7110) 7410 6510 6810 Serial # ARWM-0025
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ERROR 032

Part #	Description	Qty	Total
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Could not duplicate error 032. Retouched solder joints on the EC motor + switches. Full cal ops check + Q.C.

Technician: CAD

Date: 11/21/06