

Return & Repair Form:

Customer Information:

Company Name: Clifton Police Dept. B: _____ S: _____
 Date Received: 1-13-10 Date Given to Service: _____
 Carrier: RedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other

Product Information:

Product: 7410 7110 6510-6810-7510 Serial #: AR WM-0025
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: CUVETTE

Part Number	Description	Qty	Total Cost
<u>6809500</u>	<u>CUVETTE</u>	<u>1</u>	<u>N/C WARR</u>
<u>6809512</u>	<u>13mm BRASS</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

NEW CUVETTE
CAL W/ QC : Ops Check
JOE TALK WITH TR ARROYO #291 and had
the 7110 memory download.

Service Technician: _____

Date: 1/15/2010