

RETURN AND REPAIR FORM



Customer Information

Company Name: Clifton P.D., Clifton, NJ

Date Received: 01/07/14

Date given to service: 01/07/14

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WM-0025

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: May 2013 (59806)

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter
 Mouthpieces
 Dry Gas

Regulator
 9510 Stylus
 Other (please specify) _____

Printer Ribbon
 9510 Top Cover

Printer Paper
 Carrying Case

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	
MP Labor	Labor	1	

Repair Notes: Re-adjusted the configurations on the EC sensor readings and cleaned debri out from the breath inlet valve.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 01/10/14