

Return & Repair Form:

Customer Information:

Company Name: LITTLE FALLS POLICE B: _____ S: _____
 Date Received: 12-2-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: ARWM - 0027
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -

ACCESSORIES:

- 110 V A/C Adapter
- Regulator
- Mag Card Rdr# _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: HIGH IR RESULTS

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>6809458</u>	<u>DETECTOR</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACE FC / INSPECT COVERTE + FILTER + DETECTOR ALL LOOK
OK - REPLACED DETECTOR AS P.M.
CAL W/ QC & OPS CHECKS

Service Technician: _____

Date: 1203-08