

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Pleasantville Police Dept. Cust # _____
 Date of Return: 1/18/07 Carrier: FEDEX Method: Standard

Product Information:

Product : 7110 7410 6510 6810 Serial # ARWM-0028
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 Probe Serial # _____
 OTHER : _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: repair per attached

| Part # | Description | Qty | Total |
|----------------|------------------|-----------|-------|
| <u>6208455</u> | <u>Fuel Cell</u> | <u>1</u> | |
| <u>MPLABOR</u> | <u>LABOR</u> | <u>.5</u> | |
| <u>MRCAL71</u> | <u>7110 CAL</u> | <u>1</u> | |
| | | | |
| | | | |

Repair Notes: Replaced questionable F/C. Full cal
ops check + O.C.

Technician: CRD

Date: 1/19/07