

# Return & Repair Form:

## Customer Information:

Company Name: Pleasant PD. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 12-7-2010 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo  
 Serial #: AR WM-0028  
 Printer Ser #: AR  
 Sim Ser #: \_\_\_\_\_  
 Whole Instrument Top 1/2  
 Other \_\_\_\_\_ Probe Ser # DD P  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

## Test #

Reason for Return: PCB

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C W</u>
<u>MP CAL 21</u>	<u>CAL</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

## Repair Notes:

FC # AR AM-0425  
Install new PCB  
cal w/DC; ops check

Service Technician: \_\_\_\_\_

Date: 12-8-2010