

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER City of Wildwood Police Cust # \_\_\_\_\_  
 Date of Return: 3/9/17 Carrier: FED EX Method: Standard.

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARWM-0032  
 Description: A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: FC Problems

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLAB012</u>	<u>LABOR</u>	<u>.5</u>	
<u>MDCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced questionable F/C. Full cal  
ops check + Q.C.

Technician: CRD

Date: 3/12/07