

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Egg Harbor City P.D., Egg Harbor City NJ

Date Received: 02/28/12 Date given to service: 02/28/12

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WM-0033
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FUEL CELL</u>	<u>1</u>	<u>N/C W</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C W</u>
<u>mp cal 71</u>	<u>calibrate</u>	<u>1</u>	<u>N/C W</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Removed busted Fuel Cell AR YD-1384, installed Fuel Cell AR CK-1078.

cal w/pc, eps check

Service Technician: [Signature]

Date: 3/7/12