

N/S

Dräger safety

Internal Use Only

Return & Repair Form

Customer Information:

CUSTOMER Gallopway Township Police Dept Cust # 601127
 Date of Return: 7/11/07 Carrier: Fedex Method: P1

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWM-0035
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): POWER CORD

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: FAULTY EC?

Part #	Description	Qty	Total
<u>6808455</u>	<u>FUEL CELL</u>	<u>1</u>	<u>warranty</u>
<u>MPCAC71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C</u>

Repair Notes: NEW FC = ARXN-0407 - PROBLEM WITH FUEL CELL
NOT DUPLICATED - SEE EXAMPLE #17 REPLACED CELL
BASED ON FIELD INFO: DATA / FULL QC & OPS CHECKS

Technician: [Signature]

Date: 07-12-2007