

# Return & Repair Form:

## Customer Information:

Company Name: Galloway Twp B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 7-20-09 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: ARWM-0035  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: PCB

Part Number	Description	Qty	Total Cost
<u>8315075</u>	<u>PCB</u>	<u>1</u>	<u>N/C W/ARR</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W/ARR</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C W/ARR</u>
<u>MP LABOR</u>		<u>1 HR</u>	<u>N/C W/ARR</u>

### Repair Notes:

Replace PCB, FC  
Cal w/Cal; Ops Check

Service Technician: js

Date: 7-23-2009