

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Galleway Twp P.D. Galleway NJ

Date Received: 08/02/11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WM-0035

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>n/c w</u>
<u>6808374</u>	<u>Motor</u>	<u>1</u>	<u>n/c w</u>
<u>mp cal 71</u>	<u>Calibrate</u>	<u>1</u>	<u>n/c w</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>n/c w</u>

Repair Notes: New Fuel Cell # ARCA-0618

Motor made a greating noise, replaced motor

Cal w/De! ops check

Service Technician:

Date: 8-5-11