

Return & Repair Form

Customer Information:

CUSTOMER Cape May PD. Cust # _____
 Date of Return: 2/9/07 Carrier: FedEx Method: Express

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWM-0040
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Unit won't come out of "warm up"

Part #	Description	Qty	Total
<u>12111</u>	<u>Thermal Fuse</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>1</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Thermal Fuse was blown. Unknown cause
Replaced thermal fuse. Full cal ops check
+ O.F.

Technician: CKD

Date: 2/12/07