

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER City of Cape May Cust # _____
 Date of Return: 6-13-7 Carrier: FED EX Method: Priority

Product Information:

Product: 7110 7410 6510 6810 Serial # ARWM-0040
Description: A B PLUS SCREENER Printer Serial # _____
Whole Inst. Top Half Simulator Serial # _____
OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: Problem w/ SOLN - CHANGE

Part #	Description	Qty	Total
<u>8315075</u>	<u>Mother Board</u>	<u>1</u>	
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>1</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced Faulty M/B. Replaced questionable
FC. Full cal ops check + QC.

Technician: CRD

Date: 6/18/07