

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Maywood PD

Date Received: 7-21-11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR W M - 01075
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

- 110V A/C Adapter
- Regulator
- Printer Ribbon
- Printer Paper
- Mouthpieces
- 9510 Stylus
- 9510 Top Cover
- Carrying Case
- Dry Gas
- Other (please specify) _____

Repair Information

Test#: _____

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C W</u>
<u>MP CAL 71</u>	<u>Calibrations</u>	<u>1</u>	<u>N/C W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Upgraded Firmware N1311, Installed new Fuel Cell # APCA-0210

Cal W/DC & ops check

Service Technician: [Signature] Date: 7-26-11