

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Avon By The Sea P.D., Avon By The Sea, NJ

Date Received: 08/27/15 Date given to service: 08/27/15

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR WM-0077
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: November 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	2	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
6809458	IR Detector	1	NC/W

Repair Notes: Replaced defective IR Detector, fuel cell, and spacer plates.

IR Detector was changed due to Error 23.

Fuel Cell = ARHE-1545

CAL W/QC AND OPS CHECK

Service Technician BC Date: 09/03/15