

Return & Repair Form:

Customer Information:

Company Name: Ramsey Police Dept. B: _____ S: _____

Date Received: 11-24-09 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510
9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo

Serial #: AR WM - 1084

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: _____

Other _____

Probe Ser # DD P -

ACCESSORIES:

110 V A/C Adapter Regulator Mag Card Rdr# _____

Printer Paper Printer Ribbon Casio # _____

Mouthpieces Carrying Case Dry Gas

Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>0808455</u>	<u>FC</u>	<u>1</u>	<u>n/c warr</u>
<u>mt cal 71</u>	<u>cal</u>	<u>1</u>	<u>n/c warr</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>n/c warr</u>

Repair Notes:

FC ARWS-0037 replaced with FC ARAS-0077

cal w/oc & ops check

Service Technician: js

Date: 12-8-09