

**Return & Repair Form:**

**Customer Information:**

Company Name: Ramsy ID B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 2-15-11 Date Given to Service: \_\_\_\_\_  
 Carrier: RedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY Other

**Product Information:**

Product: 7410-~~7110~~-6510-6810-7510 Serial #: ARWM-0084  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

**ACCESSORIES:**

110 V A/C Adapter       Regulator       Mag Card Rdr# \_\_\_\_\_  
 Printer Paper       Printer Ribbon       Casio # \_\_\_\_\_  
 Mouthpieces       Carrying Case       Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

**Test #**

Reason for Return:

Part Number	Description	Qty	Total Cost
<u>mp cal 71</u>	<u>Cal</u>	<u>1</u>	
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	

Repair Notes:

There was no warranty work done, there were spots on the cuvette mirrors, mirrors were cleaned and unit calibrated.

Cal w/ QC & ops check

Service Technician: [Signature]

Date: 2/17/2011