

Return & Repair Form:

Customer Information:

Company Name: River Edge B: _____ S: _____
 Date Received: 1-20-09 Date Given to Service: _____
 Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR WM-0088
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

| Reason for Return: <u>FC</u> | | | |
|------------------------------|-------------|-----------|---------------|
| Part Number | Description | Qty | Total Cost |
| <u>6808455</u> | <u>FC</u> | <u>1</u> | <u>MCWARR</u> |
| <u>MPCAL 71</u> | <u>CAL</u> | <u>1</u> | <u>MCWARR</u> |
| | | | |
| | | | |
| | | | |
| | | | |
| <u>MPLABOR</u> | | <u>.5</u> | <u>MCWARR</u> |

Repair Notes:

REPLACE FC
CAL w/ QC 5 OPS CHECKS

Service Technician: [Signature]

Date: 01-21-09