

Return & Repair Form:

Customer Information:

Company Name: North Carolina Police B: _____ S: _____
 Date Received: 9-4-09 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR X A - 0042
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: HORN DOES NOT WORK.

| Part Number | Description | Qty | Total Cost |
|------------------|--------------|-----------|-----------------|
| <u>680 8755</u> | <u>FC</u> | <u>1</u> | <u>N/E WARR</u> |
| <u>MPCal 71</u> | <u>Cal</u> | <u>1</u> | <u>N/E WARR</u> |
| <u>MIP Labor</u> | <u>Labor</u> | <u>.5</u> | <u>N/E WARR</u> |

Repair Notes:

Replace FC
Cal w/QC & Ops check

Service Technician: JS

Date: 09-06-09