

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Atlantic City PD

Date Received: 3-26-12

Date given to service: 3-26-12

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: ARXB-0064

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
<u>mp Cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>N/C W</u>
<u>680 8455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C W</u>
<u>mp Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Replace Fuel cell, would not pass
calibration tests new Fuel cell ARCM-2323.

Cal w/ ac and ops check

Service Technician: [Signature]

Date: 3-29-12