

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Kearny P.D., Kearny, NJ

Date Received: 03/17/15

Date given to service: 03/17/15

Carrier: ☒ FedEx

☐ UPS

☐ USPS

Shipping Method:

☐ GRD

☐ NDA-PRI

☒ 3DAY

☐ 2DAY

☐ NDA-STD

Product:

☐ 6510

☐ 6810

☒ 7110

☐ 8610

☐ 7510

☐ DT5000

☐ 7410 Upper-half

☐ 9510

☐ 7410 Whole

Serial Number: AR XB-0065

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Mouthpieces

☐ Dry Gas

☐ Regulator

☐ 9510 Stylus

☐ Other (please specify) _____

☐ Printer Ribbon

☐ 9510 Top Cover

☐ Printer Paper

☐ Carrying Case

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARFJ-0670

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/24/15