

# RETURN AND REPAIR FORM

**Dräger**

## Customer Information

B:

S:

Company Name: South Orange P.D., South Orange, NJ

Date Received: 02/04/15

Date given to service: 02/04/15

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR XB-0068

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

| Part Number | Description   | Qty | Total Cost |
|-------------|---------------|-----|------------|
| 4414161     | Calibration   | 1   | NC/W       |
| 4414166     | Labor         | .5  | NC/W       |
| 6808455     | Fuel Cell     | 1   | NC/W       |
| 6808486     | Spacer Plates | 2   | NC/W       |
|             |               |     |            |
|             |               |     |            |
|             |               |     |            |
|             |               |     |            |

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARFJ-0864

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 02/06/15