

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: City of Paterson, Paterson, NJ

Date Received: 03/03/15

Date given to service: 03/03/15

Carrier: ☐ FedEx ☒ UPS ☐ USPS Shipping Method: ☒ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR XB-0073

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: May 2013

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	2	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	
8312820	Pump	1	

Repair Notes: Replaced defective pump, fuel cell, and spacer plates.

IRV was too high and readjusted to the proper settings.

Fuel Cell = ARFJ-0646

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 03/23/15