

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Mahwah P.D. Mahwah, N.J.

Date Received: 09/27/11 Date given to service: _____

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole
 Serial Number: AR XB-0078
 Printer Serial#: AR _____
 Sim Serial#: DD _____
 Probe Serial#: DD _____

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
mp cal 71	calibrator	1	w/c w
mp labor	labor	.5	w/c w

Repair Notes: Unit would not connect to the modem, I changed two capacitors located in positions C-156 and C-167.

Cal w/DC & ops check

Service Technician: [Signature] Date: 9-30-11