

Return & Repair Form:

**Customer Information:**

Company Name: Copdensburg PD B: \_\_\_\_\_ S: \_\_\_\_\_

Date Received: 09-18-89 Date Given to Service: \_\_\_\_\_

Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

**Product Information:**

Product: 7410-7110-6510-6810-7510  
9510- Drug Tester 5000

Serial #: ARXB - 0080

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top ½

Sim Ser #: \_\_\_\_\_

Other \_\_\_\_\_

Probe Ser # DD P -

**ACCESSORIES:**

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# \_\_\_\_\_

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # \_\_\_\_\_

☐ Mouthpieces

☐ Carrying Case

☐ Dry Gas

Other (Please Specify) \_\_\_\_\_

Warranty Exp. Date \_\_\_\_\_

**Repair Information:**

Test #

Reason for Return: \_\_\_\_\_

| Part Number      | Description  | Qty       | Total Cost      |
|------------------|--------------|-----------|-----------------|
| <u>mp Cal 71</u> | <u>Cal</u>   | <u>1</u>  | <u>N/K Warr</u> |
|                  |              |           |                 |
|                  |              |           |                 |
|                  |              |           |                 |
|                  |              |           |                 |
|                  |              |           |                 |
|                  |              |           |                 |
| <u>mp Labor</u>  | <u>Labor</u> | <u>.5</u> | <u>N/K Warr</u> |

Repair Notes:

Filter was loose, reinstalled Filter

Cal w/DC : Ops Check

Service Technician: ps

Date: 09-21-89