

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Old Tappan P.D., Old Tappan, NJ

Date Received: 11/30/15

Date given to service: 11/30/15

Carrier: ☐ FedEx

☒ UPS

☐ USPS

Shipping Method:

☒ GRD

☐ NDA-PRI

☐ 3DAY

☐ NDA-STD

☐ 2DAY

Product:

☐ 6510

☐ 6810

☒ 7110

☐ 8610

☐ 7510

☐ DT5000

☐ 7410 Upper-half

☐ 9510

☐ 7410 Whole

Serial Number: AR XB-0087

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description:

☐ A

☐ B

☐ Plus

☐ Demo

☐ Screener

☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Mouthpieces

☐ Dry Gas

☐ Regulator

☐ 9510 Stylus

☐ Other (please specify) _____

☐ Printer Ribbon

☐ 9510 Top Cover

☐ Printer Paper

☐ Carrying Case

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	1	NC/W
6809514	Breath Hose	1	NC/W

Repair Notes: Replaced defective breath hose and breath hose pcb.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 12/04/15