

Return & Repair Form:

Customer Information:

Company Name: North Wildwood P.D. B: 60561 S: _____

Date Received: 5-24-10 Date Given to Service: _____

Carrier: FedEx - UPS - DHL - USPS Method: GRD NDA - 2DY - 3DY - Other _____

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR XC -0003
9510- Drug Tester 5000

Description: A - B - Plus - Screener - Demo Printer Ser #: AR -

Whole Instrument Top ½ Sim Ser #: _____

Other _____ Probe Ser # DD _____ P -

ACCESSORIES:

☐ 110 V A/C Adapter ☐ Regulator ☐ Mag Card Rdr# _____

☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____

☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas

Other (Please Specify) _____ Warranty Exp. Date YES

Repair Information:

Test #

Reason for Return: PCB / FC

Part Number	Description	Qty	Total Cost
<u>835075</u>	<u>PCB</u>	<u>1</u>	<u>w/c w</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>w/c w</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>w/c w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>w/c w</u>

Repair Notes:

FC # AR XC 13 - OUT OF DATE, install new
FC AR AN-1823

Replaced PCB, old PCB missing some lot
in memory

cal w/DC; ops check

Service Technician: [Signature]

Date: 05-25-2010