

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Freehold Boro. P.D., Freehold Boro., NJ

Date Received: 07/29/2016

Date given to service: 07/29/2016

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR XC-0067

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	2.5	
6808455	Fuel Cell	1	
12029	Power Supply	1	

Repair Notes: Replaced defective fuel cell, 3 screws for power supply, and power supply.

Unit will not power on and the cover over the power supply was not on correctly. Three of the screws from the power supply was not on correctly. Three of the screws from the power supply are missing and one of the standoffs under power supply was sitting on top of the power supply.

Customer sent unit in for control test failures. Customer didn't report unit not having power.

Fuel Cell = ARJC-2221

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 09/16/2016