

# RETURN AND REPAIR FORM

# Dräger

## Customer Information

B:

S:

Company Name: Rumson Boro. P.D., Rumson, NJ

Date Received: 01/17/2018

Date given to service: 01/17/2018

Carrier: ☐ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR XC-0087

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

| Part Number | Description   | Qty | Total Cost |
|-------------|---------------|-----|------------|
| 4414161     | Calibration   | 1   |            |
| 4414166     | Labor         | .5  |            |
| 6808455     | Fuel Cell     | 1   |            |
| 6808486     | Spacer Plates | 2   |            |
|             |               |     |            |
|             |               |     |            |
|             |               |     |            |
|             |               |     |            |

Repair Notes: Replaced defective fuel cell and spacer plates.

FUEL CELL = ARKL-1129

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 02/14/2018