

Return & Repair Form:

Customer Information:

Company Name: Englewood B: _____ S: _____

Date Received: 1-4-11 Date Given to Service: _____

Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY Other _____

Product Information:

Product: 7410-7110-6510-6810-7510
9510- Drug Tester 5000

Serial #: AR XC - 0088

Description: A - B - Plus - Screener - Demo

Printer Ser #: AR -

Whole Instrument

Top 1/2

Sim Ser #: _____

Other _____

Probe Ser # DD _____ P -

ACCESSORIES:

☐ 110 V A/C Adapter

☐ Regulator

☐ Mag Card Rdr# _____

☐ Printer Paper

☐ Printer Ribbon

☐ Casio # _____

☐ Mouthpieces

☐ Carrying Case

☒ Dry Gas

Other (Please Specify) _____

Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FL</u>	<u>1</u>	<u>N/K W</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/K W</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/K W</u>

Repair Notes:

Unit was TEST with .1%, .04%, .08% and .12% before Cal process.

Cal w/AC; ops check

NEW FL ARBK-0780

Service Technician: [Signature]

Date: 01/12/2011