

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Belleville P.D., Belleville, NJ

Date Received: 02/25/14

Date given to service: 02/25/14

Carrier: ☒ FedEx

☐ UPS

☐ USPS

Shipping Method: ☒ GRD

☐ NDA-PRI

☐ 3DAY

☐ 2DAY

Product:

☐ 6510

☐ 6810

☒ 7110

☐ 8610

☐ 7510

☐ DT5000

☐ 7410 Upper-half

☐ 9510

☐ 7410 Whole

Serial Number: AR XC-0092

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Regulator

☐ Printer Ribbon

☐ Printer Paper

☐ Mouthpieces

☐ 9510 Stylus

☐ 9510 Top Cover

☐ Carrying Case

☐ Dry Gas

☐ Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced fuel cell, and spacer plates.

Fuel Cell = AREL-1315

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/04/14