

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Hoboken P.D. ,Hoboken , NJ

Date Received: 04/24/12

Date given to service: 04/24/12

Carrier: ☒ FedEx

☐ UPS

☐ USPS

Shipping Method:

☐ GRD

☐ NDA-PRI

☒ 3DAY

☐ 2DAY

☐ NDA-STD

Product:

☐ 6510

☐ 6810

☒ 7110

☐ 8610

☐ 7510

☐ DT5000

☐ 7410 Upper-half

☐ 9510

☐ 7410 Whole

Serial Number: AR XC-0101

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter

☐ Regulator

☐ Printer Ribbon

☐ Printer Paper

☐ Mouthpieces

☐ 9510 Stylus

☐ 9510 Top Cover

☐ Carrying Case

☐ Dry Gas

☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP CAL 71	Calibration	1	N/C W
6808455	Fuel Cell	1	N/C W
8315075	PCB (Mother Board)	1	N/C W
MP Labor	Labor	.5	N/C W

Repair Notes: Replaced defective fuel cell , Installed new fuel cell # ARCM-1764. Unit would not download Data properly, Installed a new (PCB).

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 05-17-12