

Return & Repair Form

Customer Information:

Company Name: CLOSTER P.D. B: _____ S: _____
Date Received: 3.24.09 Date Given to Service: 3.25.09
Carrier: FedEx UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other _____

Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR XC - 0102
Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
Whole Instrument Top ½ Sim Ser #: _____
Other _____ Probe Ser # DD P -
ACCESSORIES:
☐ 110 V A/C Adapter ☐ Regulator ☐ Mag Card Rdr# _____
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # _____
☐ Mouthpieces ☐ Carrying Case ☐ Dry Gas
Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: ERROR 32 - 2ND TIME

Part Number	Description	Qty	Total Cost
<u>6807819</u>	<u>SWITCH</u>	<u>1</u>	<u>N/C WARR</u>
<u>6808374</u>	<u>MOTOR</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP CAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>AMPLADOR</u>		<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACE MOTOR & SWITCH
FCIS 2D - INSTALLED 8-1-08
CAL W/ QC & OPS CHECKS

Service Technician: [Signature]
Created on 4/10/2008 7:22 AM

Date: 3.25.09