

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: East Newark P.D., East Newark, NJ

Date Received: 04/21/2016

Date given to service: 04/21/2016

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☐ GRD ☐ 3DAY ☐ 2DAY
☐ NDA-PRI ☒ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR XD-0003

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2013 (77967)

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	4	
6808455	Fuel Cell	1	
6808486	Spacer Plates	2	
8315075	Alcotest 7110 Motherboard	1	

Repair Notes: Replaced defective motherboard, fuel cell, and spacer plates.

Motherboard was changed due to linearity issues.

Fuel Cell = ARJC-1524

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 06/03/2016