

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Essex Fells P.D. , Essex Falls NJ

Date Received: 04/24/12

Date given to service: 04/24/12

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR XD-0009

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case
☐ Dry Gas ☐ Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
6808455	Fuel Cell	1	N/C W
MP CAL 71	Calibration	1	N/C W
MP Labor	Labor	.5	N/C W

Repair Notes: Replaced defective fuel cell ARZD-2352 with new fuel cell number ARCM-1586.

CAL W/ QC and OPS CHECK

Service Technician JS

Date: 05-02-12